



A multicenter evaluation of a point of care CRP Test

Volkher Scharnhorst^{a,*}, Peter G. Noordzij^b, Andrea Lutz^c, Uwe Graser^c, Daniel Püntener^c, Aitor Alquézar-Arbé^d

^a Clinical Laboratory, Catharina Hospital and Technical University Eindhoven, Eindhoven, The Netherlands

^b Department of Anesthesiology, Intensive Care and Pain Medicine, St. Antonius Hospital, Nieuwegein, The Netherlands

^c Roche Diagnostics International Ltd, Rotkreuz, Switzerland

^d Servicio de Urgencias, Hospital de la Santa Creu i Sant Pau, Barcelona, Spain

ARTICLE INFO

Keywords:

Antibiotic use
POC CRP Test
C-reactive protein
Infection
Inflammation
Point-of-care testing
Diagnostic equipment
Laboratory medicine

ABSTRACT

Background: Point-of-care (POC) C-reactive protein (CRP) testing in the primary healthcare setting is a cost-effective approach for reducing antibiotic prescriptions, but has yet to be widely adopted.

Methods: Analytical performance of the cobas CRP Test on the cobas b 101 system was evaluated at three POC sites and one reference laboratory. Within-run (repeatability), within-laboratory (intermediate precision), and between-laboratory precision (reproducibility) were assessed. Method comparison (reference test: CRPNX reagent [cobas c 501 module]) and matrix/lot-to-lot comparison experiments were conducted using prospectively collected blood samples from 217 adults (apparently healthy or with clinically relevant conditions). Usability and reliability were assessed by questionnaire and error reporting.

Results: Coefficients of variation (CV) for repeatability and intermediate precision ranged from 1.7%–4.0% and 1.9%–4.5%, respectively, for human serum pools containing CRP 4.7–350.7 mg/L; repeatability in clinical samples ranged from 1.6%–5.9% (3.3–360.3 mg/L). CVs for reproducibility ranged from 2.5%–4.0% (4.7–344.3 mg/L). CRP concentrations were comparable for capillary whole blood, serum, Li-heparin whole blood/plasma, K2 and K3 EDTA whole blood/plasma (Pearson's $r \geq 0.996$), and among three CRP Test lots ($r \geq 0.993$). Clinically relevant CRP concentrations measured with the CRP Test showed good agreement with those measured by CRPNX reagent (serum, weighted Deming regression $y = 0.97x + 0.11$; Pearson's $r \geq 0.996$). The overall mean usability score was 4.18/5 and the error rate across 9378 tests was 1.00%.

Conclusions: The cobas CRP Test on the cobas b 101 system demonstrates robust analytic performance when used by healthcare professionals in the POC setting.

1. Introduction

C-reactive protein (CRP) is a key mediator of the acute-phase response, with blood levels of CRP increasing rapidly after an inflammatory stimulus [1–3]. Therefore, changes in serum levels of CRP are a clinically useful marker of infection, inflammation, and tissue injury [4]. CRP testing in the primary setting (i.e. point-of-care [POC]) can help reduce diagnostic uncertainty by differentiating between bacterial and viral infections [5–7] and has been shown to be cost-effective for reducing inappropriate antibiotic prescriptions [8–12].

UK National Institute for Health and Care Excellence (NICE) guidelines recommend consideration of POC CRP testing for people

presenting with symptoms of lower respiratory tract infection in primary care, if pneumonia has not been diagnosed and it is unclear whether antibiotics should be prescribed [13]. Results of the CRP test can be used to guide antibiotic prescribing in people without a clinical diagnosis of pneumonia [13]. However, despite the availability of POC devices, POC CRP testing has not yet been widely adopted in primary care clinics [13–16].

The cobas b 101 POC system (Roche Diagnostics International Ltd., Rotkreuz, Switzerland) provides glycated hemoglobin (HbA_{1c}) and lipid panel tests (measurement of cholesterol, triglyceride, and high-density lipoprotein; calculation of low-density lipoprotein) for managing diabetes and dyslipidemia at point-of-need. The test options on the cobas b

Abbreviations: CRP, C-reactive protein; CV, coefficient of variation; HSP, human serum pool; NICE, National Institute for Health and Care Excellence; POC, point-of-care

* Corresponding author at: Clinical Laboratory, Catharina Hospital, PO Box 1350, 5602 ZA Eindhoven, The Netherlands.

E-mail addresses: volkher.scharnhorst@catharinaziekenhuis.nl (V. Scharnhorst), p.noordzij@antoniusziekenhuis.nl (P.G. Noordzij), andrea.lutz@roche.com (A. Lutz), uwe.graser@roche.com (U. Graser), daniel.puentener@roche.com (D. Püntener), aalquezar@sanpau.cat (A. Alquézar-Arbé).

<https://doi.org/10.1016/j.clinbiochem.2019.06.009>

Received 17 January 2019; Received in revised form 11 June 2019

Available online 21 June 2019

0009-9120/© 2019 Roche Diagnostics International Ltd. Published by Elsevier Inc. on behalf of The Canadian Society of Clinical Chemists. This is an open access article under the CC BY-NC-ND license (<http://creativecommons.org/licenses/by-nc-nd/4.0/>).

101 system have recently been expanded to include POC testing for CRP. We evaluated the analytical performance, usability and reliability of the cobas POC CRP Test (Roche Diagnostics International Ltd., Rotkreuz, Switzerland) and conducted a method comparison versus a reference comparator.

2. Methods

2.1. Study design

A multicenter evaluation of the cobas POC CRP Test was performed at three hospitals (Hospital Sant Pau, Barcelona, Spain; Catharina Ziekenhuis, Eindhoven, The Netherlands; St Antonius Ziekenhuis, Nieuwegein, The Netherlands), and one reference laboratory in Japan between May and November 2017. Each healthcare professional (nurses and physicians with limited POC device experience and no technical laboratory background) received a short training on how to obtain the sample, use the disk, and operate the device. A total of three CRP Test lots and three CRP Control lots were evaluated; each site was assigned two POC CRP Test lots and two CRP Control lots.

Within-run precision (repeatability), within-laboratory precision (intermediate precision), and between-laboratory precision (reproducibility) were assessed using six human serum pools (HSPs; Biomex GmbH [D-Heidelberg] and Kentucky Clinical Trials Laboratory [Louisville, USA]; HSP 1, < 5 mg/L [healthy]; HSP 2, ~10 mg/L [cutoff]; HSP 3, ~40 mg/L [decision]; HSP 4, ~100 mg/L [acute]; HSP 5, ~280 mg/L [acute high]; HSP 6, spiked with recombinant CRP to ~350 mg/L [acute high]) and two cobas CRP Control pools (level 1, representing a low-range sample; level 2, representing a high-range sample). Method comparison, matrix comparison, lot-to-lot comparison experiments, and an additional assessment of repeatability were performed using samples collected prospectively from adults aged ≥ 18 years who were either apparently healthy or who had clinically relevant conditions including inflammatory disorder and associated diseases (e.g. Crohn's disease, acute asthma without infection), infection (e.g. strong evidence of infection in the clinical record, sepsis) or tissue injury (e.g. contusion, fracture, open wound, surgery). Exclusion criteria were pregnancy or breastfeeding, and any individual whom the attending clinician deemed to be clinically unstable or whose condition could be compromised by blood draw. Participants at each POC site provided two capillary blood samples (one per lot tested) and a venous whole-blood sample for measurement of serum CRP; samples were not treated with an anticoagulant. Participants enrolled at POC site 1 provided additional venous whole-blood samples for measurement of CRP in EDTA (K2 and K3) and lithium (Li)-heparin whole blood/plasma. Samples were processed and tested immediately at each site; serum samples were split into three 1.5 mL aliquots, frozen at -70°C or colder, and shipped on dry ice for method comparison experiments performed at the reference laboratory.

The study protocol was approved by the relevant Institutional Review Board/Independent Ethics Committee at each study site prior to study initiation and the study was conducted in accordance with the Declaration of Helsinki [17]. Written informed consent was provided by all participants. The study was sponsored by Roche Diagnostics International Ltd.

2.2. Point-of-care CRP Test

The cobas POC CRP Test is an *in vitro* diagnostic test for the quantitative determination of CRP in human capillary whole blood and serum, and EDTA (K2/K3) and Li-heparin anticoagulated whole blood and plasma. The capillary whole-blood, serum or plasma sample is applied to the disk, which is inserted into the cobas b 101 instrument. Once the system is initiated, the liquid component of the sample is separated by centrifugation (i.e. separation from the blood cells), the isolated plasma or serum is diluted with dilution buffer, and CRP

present in the sample binds with the CRP antibody–latex conjugate. The concentration of CRP is determined by photometric measurement of the latex agglutination reaction using wavelengths of 525 and 625 nm. Blood flow within the CRP test is illustrated in Supplementary Fig. 1. All experiments with the POC CRP Test in the present study were performed according to the manufacturer's instructions.

2.3. Precision in clinical samples, whole-blood and plasma samples

Repeatability of the POC CRP Test was evaluated at a single site using EDTA K3 whole-blood and plasma samples. For each of the three reagent lots, one sample from each of the following CRP concentration categories was tested: healthy (< 5 mg/L); cutoff (~10 mg/L); decision (~40 mg/L); acute (~100 mg/L); acute high (~350 mg/L). For each sample, a series of 21 sequential measurements was performed to assess within-run precision.

2.4. Precision according to CLSI EP05-A3 guidelines

Repeatability and intermediate precision of the POC CRP Test were also assessed according to Clinical and Laboratory Institute (CLSI) EP05-A3 guidelines [18]. Samples comprised six HSPs and two CRP Controls (levels 1 and 2). For each lot ($n = 2$ lots per site), samples were measured four times daily for 21 days (two replicates of each sample for each of two runs per day). Samples were analyzed in random order per day for 21 days.

Reproducibility of the POC CRP Test was assessed at three sites over 5 days using HSPs 1–6 and CRP Controls (levels 1 and 2). Five replicates of each sample were analyzed per day over 5 days.

2.5. Sample matrix comparison

Test performance in different matrices was compared using matched serum and capillary blood samples, and matched serum and whole-blood/plasma samples containing EDTA K2, EDTA K3, and Li-heparin as anticoagulants. Matched pairs were analyzed in parallel, in a single run for each of two POC CRP Test lots, at a single site.

2.6. Lot-to-lot comparison

POC CRP Test lot-to-lot comparison was performed using capillary whole-blood/serum samples (at three sites with a total of three disk lots) or EDTA K2, EDTA K3, and Li-heparin whole-blood/plasma samples (at a single site with two disk lots). Samples were tested in a single run per reagent lot. Multiple samples were taken from the same subject and run in parallel on two different instruments, using a different disc lot for each sample.

2.7. Method comparison

Healthcare professionals determined CRP concentrations in capillary whole blood/serum and EDTA K2, EDTA K3, and Li-heparin in whole blood/plasma using the cobas POC CRP Test (cobas b 101 system; three sites); some users had limited experience of using the system previously, and some were naïve to the system. For comparison, CRP values from matched serum samples were obtained by laboratory professionals using the cobas c 501 module (Roche Diagnostics International Ltd., Rotkreuz, Switzerland) with the CRPNX reagent (reference test; Denka Seiken, Tokyo, Japan). The reference test has a measuring range of 0.1–320 mg/L and a limit of detection of 0.05 mg/L; the CV for within-run precision ranged from 0.62–0.93% across control samples. No interference of the test is shown with ascorbic acid up to concentration of 50 mg/dL; hemoglobin up to 500 mg/dL; bilirubin up to 30 mg/dL and intralipid up to 5 mg/dL.

2.8. Usability and reliability

Usability of the POC CRP Test on the cobas b 101 system was evaluated by questionnaires completed by all operators at each of the three sites. Questions related to six individual domains: general; general aspects of the software; processing of the sample; CRP disk; quality control; and cleaning/disinfection. The usability questionnaire employed a 5-point scoring system (1, very poor; 2, poor; 3, average; 4, good; 5, excellent), with mean scores calculated for each domain and overall. Reliability of the system was assessed by capturing CRP Test and instrument errors. An overall error rate was calculated based on reported CRP Test and instrument errors.

2.9. Data analysis and statistical methods

The target sample size was ≥ 60 adults per site to provide samples covering the measuring range of the POC CRP Test (3–400 mg/L) in the following prespecified distribution: < 20 mg/L ($> 40\%$ of participants); 20–100 mg/L, 100–200 mg/L, and 200–400 mg/L (each $> 10\%$ of participants). All data were evaluated and analyzed with validated statistical tools (WinCAEv and Medrio, Mannheim, Germany; SAS software, SAS Institute Inc., Cary, NC).

For precision experiments, the standard deviation (SD; for low concentration samples) and coefficient of variation (CV) were used as the measure of assay imprecision. Matrix and lot-to-lot comparisons were evaluated using Passing–Bablok analysis. Agreement between methods was evaluated using weighted Deming regression analysis. Bias was examined using Bland–Altman analysis, with relative difference calculated as $(Y-X)*100/X$ (where $X = \text{CRPNX test [serum]}$ and $Y = \text{POC CRP Test}$). Mean bias between CRP values on the different systems was calculated and the relative bias at medical decision points was determined.

3. Results

3.1. Samples

A total of 217 adult patients were recruited from semi-intensive care, cardiovascular care, ambulatory care or the emergency department, and comprised 158 males (72.8%) and 59 females (27.2%). Median age (range) was 68 (range, 22–88) years. Participants with valid measurements ($n = 201$) were categorized according to clinical relevance for CRP, as follows: tissue injury ($n = 116$ [57.7%]), infection ($n = 54$ [26.9%]), inflammatory disorders and associated diseases ($n = 28$ [13.9%]) or CRP ≤ 5 mg/L ($n = 3$ [1.5%]). Samples ($n = 201$) were categorized by CRP concentration range (Supplemental Table 1).

3.2. Precision in clinical samples, whole-blood and plasma samples

Repeatability results for EDTA K3 whole-blood samples (mean CRP concentration 3.3–360.3 mg/L) and plasma samples (3.3–356.5 mg/L) are presented in Supplemental Table 2. Across sample types, CVs for repeatability ranged from 1.6% to 5.9%.

Table 1

Precision of the POC CRP Test, as assessed according to CLSI EP05-A3 guidelines. Data are presented as a range across individual lots and sites; each range is based on $n = 504$ measurements ($n = 84$ replicate measurements per sample, per lot, across three sites). See Supplementary Table 1 for results by site and by lot.

	HSP 1	HSP 2	HSP 3	HSP 4	HSP 5	HSP 6	Control level 1	Control level 2
Mean CRP, mg/L	4.7–4.8	10.3–10.6	42.3–43.2	101.5–104.1	224.0–231.5	342.0–350.7	9.6–10.1	44.1–45.8
Intermediate precision, CV (%)	3.3–4.5	2.5–3.7	2.9–4.4	1.9–3.1	2.7–3.5	2.6–3.5	3.3–5.0	3.1–4.1
Repeatability, CV (%)	3.1–4.0	1.8–3.1	2.5–3.4	1.7–2.2	2.2–3.2	2.0–2.9	2.6–4.7	2.4–3.8

CRP, C-reactive protein; HSP, human serum pool.

3.3. Precision according to CLSI EP05-A3 guidelines

Repeatability and intermediate precision are presented in Table 1 (overall) and Supplemental Table 3 (by lot and site) and were consistent across sites and lots. CVs for repeatability in HSPs ranged from 1.7% to 4.0% (site 1, 2.0%–4.0% [mean CRP range, 4.8–350.7 mg/L]; site 2, 1.7%–3.6% [mean CRP range, 4.7–344.2 mg/L]; site 3, 1.7%–3.9% [mean CRP range, 4.7–345.3 mg/L]); CVs for intermediate precision in HSPs ranged from 1.9% to 4.5% (site 1, 2.1%–4.1%; site 2, 1.9%–4.5%; site 3, 2.3%–4.2%).

The SD for reproducibility was 0.2 mg/L for the low-concentration HSP 1 sample (mean CRP concentration 4.7 mg/L). CVs for reproducibility ranged from 2.5% to 4.0% for HSPs 2–6 (mean CRP concentrations 10.4–344.3 mg/L; Table 2).

3.4. Matrix comparison

CRP concentrations measured using the different sample types correlated well with CRP concentrations measured in serum (Pearson's $r \geq 0.996$, intercept ≤ 0.32 , slope ≥ 0.96 ; Supplemental Table 4); across lots and matrices, mean bias was low, ranging from 0.02% to 0.34% for low concentration samples (0–5 mg/L), from -2.18% to 0.90% for mid-range samples (5–200 mg/L), and from -5.53% to -0.74% for higher-concentration samples (Supplemental Table 5). There was good agreement for CRP concentrations measured in serum with Li-heparin whole-blood samples (Pearson's r , 0.998; Fig. 1A) and EDTA K3 whole blood (Pearson's r , 0.998; Fig. 1B).

3.5. Lot-to-lot comparison

Correlations among different test lots were very good for whole-blood, plasma, and serum samples, with all Pearson's r values ≥ 0.993 , intercepts ≤ 0.30 , and slopes ≥ 0.96 (Fig. 2; Supplemental Table 6). Mean bias with different lots was $\leq 2.37\%$.

3.6. Method comparison

There was very good correlation between CRP values measured with the cobas POC CRP Test on the cobas b 101 POC system (all sample matrices) and values measured in matched serum samples with the CRPNX test on the cobas c 501 laboratory module (Pearson's $r \geq 0.994$, slope ≥ 0.93 , intercept ≤ 0.47 ; Supplemental Table 7). Across matrix types, lots, and sites, bias at the lower (5.0 mg/L) and upper (10.0 mg/L) medical decision points was $\leq 6.07\%$; results for capillary whole-blood and serum samples pooled by site and for anticoagulated whole-blood and plasma samples at a single site are presented in Table 3. Biases for the comparison of the POC CRP Test versus the CRPNX test (serum) are presented in Table 3. For the CRP range 3–200 mg/L, mean bias (95% CI) was -2.05% ($-10.54, 6.43$) for serum and -2.05% ($-15.64, 11.54$) for capillary whole blood. Example Bland Altman plots are shown for serum in Fig. 3A and capillary whole blood in Fig. 3B. When CRP was measured in serum samples with the POC CRP Test and CRPNX (pooled site; lot 1; $n = 130$ samples), weighted Deming regression analysis yielded a Pearson's r of 0.996 (Supplemental Fig. 2A). Weighted Deming regression analysis of the POC CRP Test

Table 2
Reproducibility of the POC CRP Test. Data shown are combined results across all three sites.

	HSP 1	HSP 2	HSP 3	HSP 4	HSP 5	HSP 6	Control level 1	Control level 2
Mean CRP, mg/L	4.7	10.4	42.5	102.1	225.6	344.3	9.7	44.7
Reproducibility, CV (%)	0.2	3.4	3.2	2.5	4.0	3.9	4.7	4.0
Site-to-site, CV (%)	0.1	1.7	0.7	0.8	0.0	0.8	0.0	0.7
Lot-to-lot, CV (%)	0.0	0.2	0.4	0.0	1.9	1.5	3.4	2.1
Day-to-day, CV (%)	0.1	0.4	1.2	0.9	0.0	1.3	0.6	1.6
Repeatability, CV (%)	0.2	2.9	2.9	2.1	3.6	3.3	3.2	2.9

CRP, C-reactive protein; CV, coefficient of variation; HSP, human serum sample pools; SD, standard deviation.

(whole capillary blood; lot 1) versus CRPNX test (serum) results, based on pooled site data for 136 samples, yielded a Pearson's r of 0.994 (Supplemental Fig. 2B).

3.7. Operator-reported usability

Overall mean score for operator-reported usability assessed across six domains was 4.18 out of 5 (i.e. between “good” and “excellent”; $n = 7$ completed questionnaires across the three sites). Mean scores by individual domain are presented in Supplemental Fig. 3. Of interest, all sites noted the rapid turnaround time for results and the ease of use of the instrument, but also the missing battery-operated modus, meaning that it was cumbersome to plug in the instruments at patient beds.

A total of 9378 POC CRP Test disks were tested during this performance evaluation and 94 errors were reported (~1% error rate). Most errors (46%) were reaction errors due to overfilling the disk where the sample is able to move out of the chamber due to backflow. The second most common error (24%) was large or insufficient sample, and most cases occurred with serum or control solution (only one error was reported with capillary whole blood). None of the reported errors were associated with wrong measurement results.

4. Discussion

Use of CRP testing at POC can help to reduce inappropriate antibiotic prescriptions [8–12]; however, widespread use in primary care settings has not yet materialized. Availability of easy-to-use and reliable

CRP testing may facilitate greater use by clinicians. In the current analyses, we show that the cobas POC CRP Test delivered precise and repeatable CRP values with clinical specimens (including samples with very high CRP values) and with HSPs measured according to CLSI EP05-A3 criteria. Notably, CVs for imprecision were well below the specifications reported in the desirable biological variation database, which state a CV for imprecision of 21.1% [19]. CRP values measured using different sample types correlated well, with low mean bias. Correlations between different test lots were very good, indicating that different lots of the POC CRP Test perform equally well. Moreover, CRP values measured in a range of sample types with the POC CRP Test demonstrated very good correlation with CRP values measured in serum samples with the CRPNX test on the cobas c 501 module; these results indicate that the POC CRP Test delivers comparable performance to laboratory CRP testing. Finally, usability of the cobas b 101 system was rated as “good” to “excellent” by operators, and reported error rates were low, indicating convenience of the system for use in the POC environment.

POC CRP testing offers several clinical and health-economic advantages. Compared with CRP testing, procalcitonin testing is more expensive in both POC and automated settings [20,21]. Notably, POC CRP testing has been shown to be cost-effective for reducing inappropriate antibiotic prescriptions in patients presenting with suspected respiratory tract infection [13,22]. A recent systematic review of randomized controlled trials and observational studies reported that POC CRP testing was associated with a significant 25% reduction in antibiotic prescribing at the index consultation relative to no POC CRP




Fig. 1. Matrix comparison: Passing–Bablok regression analysis of CRP values measured with the POC CRP Test using serum (reference) compared with (A) Li-heparin whole blood and (B) EDTA K3 whole blood (test). Data are shown for Lot 1 analyzed at POC site 1.
CRP, C-reactive protein; POC, point-of-care; WB, whole blood.




Fig. 2. Lot-to-lot comparison: Passing–Bablok regression analysis of CRP values measured with different lots of the POC CRP Test using serum (A and B) and capillary whole blood (C and D).

CRP, C-reactive protein; WB, whole blood.

testing (risk ratio 0.75; 95% confidence interval [CI], 0.67–0.83) [8]. Reductions in unnecessary antibiotic prescriptions are desirable given the growing problem of antibiotic resistance [23]. Moreover, over half of patients presenting to primary care with lower respiratory tract infections are prescribed antibiotics [24], yet there is limited evidence that the marginal benefits of these treatments outweigh the risk for harm [25,26]. Although CRP tests are available through local hospital pathology services at a lower cost, POC testing allows the findings to be discussed between patient and primary care practitioner during the consultation and may be helpful in persuading patients to agree to a non-antibiotic approach [13,27].

UK NICE guidelines recommend that POC CRP testing is considered for people presenting to primary care with symptoms of lower respiratory tract infection. The guidelines recommend that antibiotic

therapy should not be routinely offered if the CRP concentration is < 20 mg/L, that a delayed antibiotic prescription (i.e. for use at a later date, if symptoms worsen) should be considered if the CRP concentration is 20–100 mg/L, and that antibiotic therapy should be offered if the CRP concentration is ≥ 100 mg/L.

Clinical experience with existing POC CRP tests supports their use in primary care [28]. For example, Matheussen and colleagues [28] compared a CRP QuikRead POC testing device (Orion Diagnostica, Espoo, Finland) with a central laboratory method for CRP measurement in over 2900 serum samples from adult patients presenting to primary care with symptoms of lower respiratory infection. The POC device had a sensitivity of 92.2% and specificity of 99.4% at a CRP cutoff of ≥ 30 mg and was deemed a good candidate for POC CRP testing. Interestingly, the slope of the Passing–Bablok analysis suggested a 6% underestimation of CRP

Table 3
Method comparison: mean bias for CRP measured with the cobas POC CRP Test (cobas b 101 POC system) using (A) capillary whole blood and serum (data from all sites combined), (B) EDTA K2/K3 and Li-heparin in whole blood and plasma samples (data from POC site 1 only), on the CRP-X (cobas c 501 laboratory module).

A			Capillary whole blood			Serum		
Lot	CRP range/*MDP (mg/L)	n	Mean bias, % (95% CI)		n	Mean bias, % (95% CI)		
			Capillary whole blood	Serum		Capillary whole blood	Serum	
1	3–200	121	-2.05 (-15.64, 11.54)		115	-2.05 (-10.54, 6.43)		
	200–400	15	-2.41 (-16.50, 11.68)		15	-0.67 (-17.59, 16.25)		
	5*	136	2.88 (0.32, 5.44)		130	-0.32 (-2.66, 2.02)		
	10*	136	-0.64 (-1.90, 0.62)		130	-1.47 (-2.58, -0.36)		
	200–400	119	-2.39 (-15.27, 10.48)		116	-1.44 (-12.19, 9.32)		
2	3–200	17	-4.43 (-11.27, 2.40)		16	-1.34 (-11.53, 8.85)		
	200–400	136	2.44 (-0.16, 5.04)		132	2.89 (0.61, 5.18)		
	5*	136	-0.99 (-2.24, 0.27)		132	-0.04 (-1.13, 1.05)		
	10*	116	-3.26 (-13.06, 6.54)		112	-1.43 (-10.86, 8.01)		
	200–400	12	-6.38 (-18.63, 5.88)		11	-2.63 (-8.29, 3.04)		
3	3–200	128	-1.38 (-3.88, 1.12)		123	2.26 (0.31, 4.21)		
	200–400	128	-2.81 (-4.03, -1.58)		123	-0.22 (-1.16, 0.72)		
	5*							
	10*							

B			Plasma			Li-heparin												
Lot	CRP range/*MDP (mg/L)	Whole blood	EDTA K2		EDTA K3		Li-heparin											
			n	Mean bias, 95% CI (%)	n	Mean bias, 95% CI (%)	n	Mean bias, 95% CI (%)										
1	3–200	62	-1.79 (-14.11, 10.54)		62	-4.05 (-13.47, 5.37)		62	-2.58 (-12.28, 7.11)		58	-3.24 (-12.72, 6.24)		56	-4.68 (-14.04, 4.68)		53	-2.72 (-14.55, 9.10)
	200–400	10	-3.02 (-10.43, 4.40)		10	-5.22 (-14.43, 3.99)		10	-1.55 (-9.99, 6.88)		10	-2.96 (-17.75, 11.83)		10	-5.14 (-13.07, 2.79)		10	-2.87 (-8.33, 2.59)
	5*	72	1.55 (-2.67, 5.77)		72	-0.89 (-4.88, 3.11)		72	0.30 (-2.42, 3.02)		68	-0.97 (-5.01, 3.08)		66	-2.33 (-6.16, 1.49)		63	1.23 (-3.95, 6.41)
	10*	72	-1.02 (-2.95, 0.92)		72	-3.32 (-5.09, -1.55)		72	-1.72 (-3.15, -0.29)		68	-2.64 (-4.48, -0.80)		66	-4.09 (-5.78, -2.39)		63	-1.61 (-3.94, 0.72)
	200–400	62	-2.50 (-14.06, 9.06)		62	-4.38 (-17.06, 8.30)		62	-2.30 (-15.59, 11.00)		58	-3.17 (-13.92, 7.59)		57	-4.85 (-16.08, 6.38)		53	-2.14 (-11.89, 7.61)
2	3–200	10	-3.24 (-11.02, 4.54)		10	-4.38 (-12.65, 3.90)		10	-3.63 (-15.66, 8.40)		10	-5.31 (-15.03, 4.42)		10	-3.31 (-10.16, 3.54)		10	-3.51 (-9.75, 2.73)
	200–400	72	1.83 (-1.77, 5.44)		72	-0.24 (-3.91, 3.44)		72	2.98 (0.04, 5.91)		68	2.97 (-0.95, 6.88)		67	-0.15 (-3.77, 3.47)		63	1.36 (-3.10, 5.82)
	5*	72	-1.40 (-3.09, 0.28)		72	-3.26 (-4.96, -1.56)		72	-1.00 (-2.51, 0.52)		68	-1.76 (-3.55, 0.03)		67	-3.40 (-5.07, -1.73)		63	-1.29 (-3.27, 0.70)
	10*																	

CI, confidence interval; CRP, C-reactive protein; MDP, medical decision point; PL, plasma; WB, whole blood.




Fig. 3. Method comparison: Bland-Altman analysis of CRP values measured with the cobas POC CRP Test (Lot 1) on the cobas b 101 system using (A) serum (B) capillary whole blood, compared with values measured using matched serum samples on the cobas c 501 laboratory module with the CRPNX test. Data are for all sites combined. Relative difference (%) calculated as $(Y-X) \times 100/X$, where X = CRPNX test (serum) and Y = POC CRP Test. CRP, C-reactive protein.

levels by the POC device (Passing–Bablok slope 0.94; 95% CI, 0.93–0.95) [28]. For comparison, in the present study there was strong agreement between the POC CRP Test and the central laboratory reference test (CRPNX reagent); for serum samples, the slope of the weighted Deming regression was 0.97 and the upper 95% CI ranged from 0.98 to 0.99 across lots, indicating a 3% underestimation of CRP concentrations.

Strengths of the present study include the use of clinical samples derived from prospectively enrolled individuals, and that precision was measured according to CLSI EP05-A3 guidelines [18]. Importantly, the method comparison was performed by healthcare professionals in the POC environment and laboratory operators at a reference laboratory (i.e. relevant user groups), and used a relevant reference test (CRPNX). A reference range for CRP specific to the cobas POC CRP Test was not determined in the present study; therefore, further studies are warranted to define a reference range and explore the utility of this POC CRP Test in clinical populations. As this study was conducted in hospital settings, additional studies conducted in a primary care population would be beneficial to validate our findings and further examine the utility of the POC CRP Test in clinical populations. Use of the POC CRP Test would be expected to have equivalent benefit in clinical populations, particularly since all healthcare professionals are able to learn to use the device after a short training session.

In conclusion, these findings indicate that healthcare professionals can obtain precise and reproducible CRP values with the cobas POC CRP Test that show very good correlation with laboratory measurements. Importantly, operators considered the system convenient for use in the POC environment.

Acknowledgments

This study was funded by Roche Diagnostics International Ltd. (Rotkreuz, Switzerland). Medical writing assistance was provided by David Evans, PhD, and Louise Kelly, BSc, (Gardiner-Caldwell Communications, Macclesfield, UK), and was funded by Roche Diagnostics International Ltd. (Rotkreuz, Switzerland).

The funding organization played no role in the study design; in the

collection, analysis, and interpretation of data; in the writing of the report; or in the decision to submit the report for publication. All authors have accepted responsibility for the entire content of this submitted manuscript and approved submission.

Andrea Lutz, Uwe Graser and Daniel Püntener are employees of, and own stocks/shares, in Roche Diagnostics International Ltd. Dr. Alquézar-Arbé received an honorarium from Roche Diagnostics for his participation as investigator of the study. Dr. Noordzij received an honorarium from Roche for an IL-6 Advisory Board on September 20–21, 2016 in Frankfurt, Germany.

Appendix A. Supplementary data

Supplementary data to this article can be found online at <https://doi.org/10.1016/j.clinbiochem.2019.06.009>.

References

- [1] S. Black, I. Kushner, D. Samols, C-reactive protein, *J. Biol. Chem.* 279 (2004) 48487–48490.
- [2] W. Ansar, S. Ghosh, C-reactive protein and the biology of disease, *Immunol. Res.* 56 (2013) 131–142.
- [3] C. Gabay, I. Kushner, Acute-phase proteins and other systemic responses to inflammation, *N. Engl. J. Med.* 340 (1999) 448–454.
- [4] P. Póvoa, L. Coelho, E. Almeida, A. Fernandes, R. Mealha, P. Moreira, et al., C-reactive protein as a marker of infection in critically ill patients, *Clin. Microbiol. Infect.* 11 (2005) 101–108.
- [5] J.P. Haran, F.L. Beaudoin, S. Suner, S. Lu, C-reactive protein as predictor of bacterial infection among patients with an influenza-like illness, *Am. J. Emerg. Med.* 31 (2013) 137–144.
- [6] R.M. Hopstaken, J.W. Muris, J.A. Knottnerus, A.D. Kester, P.E. Rinkens, G.J. Dinant, Contributions of symptoms, signs, erythrocyte sedimentation rate, and C-reactive protein to a diagnosis of pneumonia in acute lower respiratory tract infection, *Br. J. Gen. Pract.* 53 (2003) 358–364.
- [7] S.F. van Vugt, B.D.L. Broekhuizen, C. Lammens, N.P.A. Zuihthoff, P.A. de Jong, S. Coenen, et al., Use of serum C reactive protein and procalcitonin concentrations in addition to symptoms and signs to predict pneumonia in patients presenting to primary care with acute cough: diagnostic study, *BMJ* 346 (2013) f2450.
- [8] Y. Huang, R. Chen, T. Wu, X. Wei, A. Guo, Association between point-of-care CRP testing and antibiotic prescribing in respiratory tract infections: a systematic review and meta-analysis of primary care studies, *Br. J. Gen. Pract.* 63 (2013) e787–e794.
- [9] R. Aabenhus, J.U. Jensen, K.J. Jørgensen, A. Hróbjartsson, L. Bjerrum, Biomarkers as point-of-care tests to guide prescription of antibiotics in patients with acute

- respiratory infections in primary care, *Cochrane Database Syst. Rev.* 11 (2014) CD010130.
- [10] P. Little, B. Stuart, N. Francis, E. Douglas, S. Tonkin-Crine, S. Anthierens, et al., Effects of internet-based training on antibiotic prescribing rates for acute respiratory-tract infections: a multinational, cluster, randomized, factorial, controlled trial, *Lancet* 82 (2013) 1175–1182.
- [11] R. Oppong, M. Jit, R.D. Smith, C.C. Butler, H. Melbye, S. Mølstad, et al., Cost-effectiveness of point-of-care C-reactive protein testing to inform antibiotic prescribing decisions, *Br. J. Gen. Pract.* 63 (2013) e465–e471.
- [12] J.W.L. Cals, A.J. Ament, K. Hood, C.C. Butler, R.M. Hopstaken, G.F. Wassink, et al., C-reactive protein point-of-care testing and physician communication skills training for lower respiratory tract infections in general practice: economic evaluation of a cluster randomized trial, *J. Eval. Clin. Pract.* 17 (2011) 1059–1069.
- [13] National Clinical Guideline Centre, *Pneumonia: Diagnosis and management of community- and hospital-acquired pneumonia in adults*, Clin. Guidel. 191 (2014) 3 December. Available at: <https://www.nice.org.uk/guidance/cg191/evidence/full-guideline-193389085>, Accessed date: 10 October 2018.
- [14] M.C. Minnaard, A.C. van de Pol, B.D. Broekhuizen, T.J. Verheij, R.M. Hopstaken, S. van Delft, Analytical performance, agreement and user-friendliness of five C-reactive protein point-of-care tests, *Clin. Lab. Invest.* 73 (2013) 627–634.
- [15] N. Brouwer, J. van Pelt, Validation and evaluation of eight commercially available point-of-care CRP methods, *Clin. Chim. Acta* 439 (2015) 195–201.
- [16] [No authors listed], Point-of-care CRP testing in the diagnosis of pneumonia in adults, *Drug Ther. Bull.* 54 (2016) 117–120.
- [17] World Medical Association, *WMA Declaration of Helsinki – Ethical Principles for Medical Research Involving Human Subjects*. 9 July 2018, Available at: <https://www.wma.net/policies-post/wma-declaration-of-helsinki-ethical-principles-for-medical-research-involving-human-subjects/>, Accessed date: 10 October 2018.
- [18] Clinical and Laboratory Standards Institute, *CLSI EP05-A3. Evaluation of Precision of Quantitative Measurement Procedures*, 3rd ed, NCCLS, Wayne, PA, 2014.
- [19] J. Westgard, *Quality Requirements: Desirable Specifications for Total Error, Imprecision and Bias, Derived from Intra- and Inter-Individual Biologic Variation*. Updated 2014, Available at: <https://www.westgard.com/biodatabase1.htm>, Accessed date: 10 October 2018.
- [20] F. Khorvash, F. Abdi, K. Dialami, Can serum procalcitonin and C-reactive protein as nosocomial infection markers in hospitalized patients without localizing signs? *J. Res. Med. Sci.* 16 (2011) 1280–1285.
- [21] B. Andreola, S. Bressan, S. Callegaro, Procalcitonin and C-reactive protein as diagnostic markers of severe bacterial infections in febrile infants and children in the emergency department, *Pediatr. Infect. Dis. J.* 26 (2007) 672–677.
- [22] R. Hunter, Cost-effectiveness of point-of-care C-reactive protein tests for respiratory tract infection in primary care in England, *Adv. Ther.* 32 (2015) 69–85.
- [23] European Centre for Disease Prevention and Control (ECDC). *Surveillance of antimicrobial resistance in Europe 2016, Annual Report of the European Antimicrobial Resistance Surveillance Network (EARS-Net)*, ECDC, Stockholm, 2017.
- [24] C.C. Butler, K. Hood, T. Verheij, P. Little, H. Melbye, J. Nuttall, et al., Variation in antibiotic prescribing and its impact on recovery in patients with acute cough in primary care: prospective study in 13 countries, *BMJ* 338 (2009) b2242.
- [25] S.M. Smith, J. Smucny, T. Fahey, Antibiotics for acute bronchitis, *JAMA* 312 (2014) 2678–2679.
- [26] P. Little, B. Stuart, M. Moore, S. Coenen, C.C. Butler, M. Godycki-Cwirko, et al., GRACE consortium. Amoxicillin for acute lower-respiratory-tract infection in primary care when pneumonia is not suspected: a 12-country, randomised, placebo-controlled trial, *Lancet Infect. Dis.* 13 (2013) 123–129.
- [27] J.R. Huddy, M.Z. Ni, J. Barlow, A. Majeed, G.B. Hanna, Point-of-care C reactive protein for the diagnosis of lower respiratory tract infection in NHS primary care: a qualitative study of barriers and facilitators to adoption, *BMJ Open* 6 (2016) e009959.
- [28] V. Matheussen, V. Van Hoof, K. Loens, C. Lammens, A. Vanderstraeten, S. Coenen, et al., GRACE consortium. Analytical performance of a platform for point-of-care CRP testing in adults consulting for lower respiratory tract infection in primary care, *Eur. J. Clin. Microbiol. Infect. Dis.* (2018), <https://doi.org/10.1007/s10096-018-3253-3> May 10. (Epub ahead of print).