

HISTORICAL DEVELOPMENT OF RAPID METHODS
AND AUTOMATION IN MICROBIOLOGY¹

DANIEL Y.C. FUNG

*Department of Animal Sciences and Industry
Kansas State University
Manhattan, KS 66506-1600*

Accepted for Publication January 2, 1992

ABSTRACT

Rapid methods and automation in microbiology comprise a dynamic area of research and study. The field encompasses a wide variety of subjects in theoretical and applied microbiology. This article traces the major developments of this field since the late 1960s, detailed with dates and locations of major international meetings on the subject as well as listing of major books, proceedings, and articles concerning this topic. The article should be of interest to all scientists who want to know the key developments, current status, and possible future trends of the field.

HISTORICAL DEVELOPMENT OF RAPID METHODS
AND AUTOMATION IN MICROBIOLOGY

Rapid methods and automation in microbiology comprise a relatively new field of study in applied microbiology. This field attempts to utilize microbiological, chemical, biochemical, biophysical, immunological, and serological methods for the study of improved methods for the isolation, early detection, characterization, and enumeration of microorganisms and their products in clinical, food, industrial, and environmental samples. The field is dynamic and has gained momentum in the past 10 years.

Medical microbiologists started to be involved in rapid methods around the early 1970s. In the 80s there was a surge of use of diagnostic test kits for microbiology

¹This material is based upon work supported by the Cooperative State Research Service, U.S.D.A., under Agreement No. 89-34187-4511. Part of this paper was presented at the 105th AOAC Annual International Meeting and Exposition, August 12-15, 1991, Phoenix, AZ. Contribution No. 92-332-J, Kansas Agricultural Experiment Station, Manhattan, KS 66506. Copies of Author's papers cited in this review are available upon request.

and immunology in the medical areas. Food microbiologists are lagging about 10 years behind medical microbiologists in this regard. However, in the past 5 years, much interest has been expressed internationally by food microbiologists in adapting new and rapid methods. Surely, the field will expand considerably in the next 10 years in all areas of microbiology. With advanced technologies in computers and instrumentation, we should expect much activity in the entire field.

At the dawn of the development of bacteriology and microbiology, scientists must have slowly tried to improve efficiency in the operation of bacteriological procedures. A look at older bacteriology text books and laboratory manuals will reveal some form of improved operations for handling large numbers of culture tubes, test tubes, plates, and inoculation procedures.

Although many handbooks and manuals contained descriptions of improved methods in bacteriology and microbiology, the most important book dealing with *Miniaturization of Microbiological Methods* was written by Paul A. Hartman (1968). The cover page of the book reads "This book provides a comprehensive compilation (over 1,200 citations) of techniques for the cultivation of bacteria, fungi, protozoa, and other plants and animal cells. Emphasis is given to techniques made more rapid, convenient, or reliable than conventional laboratory methods by miniaturization." Another very useful book was compiled by V.B.D. Skerman (1969) on *Abstracts of Microbiological Methods* with almost 900 pages of valuable information on methodologies in applied microbiology. These two books are classics and should be read by all students of rapid methods and automation in microbiology.

In the early 1970's, several interesting articles appeared concerning rapid methods and automation in microbiology. Goldschmidt (1970) had a chapter entitled "Instrumentation for Microbiology: Horizons Unlimited" in the book *Rapid Diagnosis Methods in Medical Microbiology* by Graber (1970). Richardson (1972) had an article on "Automation in the Dairy Laboratory" in the *Journal of Milk and Food Technology* and Trotman (1973) wrote an article on "The Philosophy of the Application of Automatic Methods to Hospital Diagnostic Bacteriology" in *Biochemical Engineering*.

One of the earliest symposia on this topic was organized by Daniel Amsterdam in 1971 at the Annual Meeting of the American Society for Microbiology in Minneapolis, Minnesota with the title "Rapid Methods for Detection and Characterization of Microorganisms." Daniel Y.C. Fung (1971) and Millicent C. Goldschmidt (1971), the editor and associate editor of the *Journal of Rapid Methods and Automation in Microbiology*, were invited speakers at that historical symposium.

Although many smaller meetings, seminars, and symposia were held concerning rapid methods, the key identifiable start of the field was the first *International Symposium on Rapid Methods and Automation in Microbiology* held in Stockholm, Sweden, under the chairmanship of Carl-Goran Heden in 1973. There were about

500 to 600 people at that meeting. Subsequently, symposia of this series were held in Cambridge, UK (1976), Washington, DC (1981), Berlin (1984), Florence, Italy (1987), and Helsinki (1990). The seventh symposium will be held in London, UK, in 1993. The meeting has grown in size to a little more than 1,000 participants in recent years. Also the name has been changed to *International Congress on Rapid Methods and Automation in Microbiology and Immunology*. This author has had the honor to be plenary session speaker, symposium organizer, and paper presenter in all these international symposia since 1973.

The proceedings of these meetings were published by Heden and Illeni (1975a,b), Newsom (1976), Tilton (1982), Habermehl (1984), Balows *et al.* (1989), and Vaheri *et al.* (1991). Of these proceedings, the most valuable ones are the two books edited by Heden and Illeni: *Automation in Microbiology and Immunology* (1975a) and *New Approaches to the Identification of Microorganisms* (1975b). These two books, although outdated by now, contain the basic approaches and philosophies of the field of rapid methods. Subsequent volumes of the symposium series provided updated information and newer developments. All these books are valuable resources in the field.

In between these major meetings, international conferences also were held on similar topics in Kiel, Germany (1974); Ottawa, Canada (1975); Dallas, Texas (1978); Liblice Castle, Czechoslovakia (1980); Ploufragan, France (1983); Lille, France (1983); Anchorage, Alaska (1986); Taipei, Taiwan (1987); Nancy, France (1988); Singapore (1989); and other places. Some of these proceedings were published: Tolle and Heesch (1974), Sharpe and Clark (1978), Schindler (1980), Lahellec (1983), LaClerc (1983), Kramer and Liston (1987), Hung (1987), and Galteau *et al.* (1989).

Organizations such as *American Society for Microbiology*, *Institute of Food Technologists*, *Biodeterioration Society*, etc., occasionally organized symposia on this topic through the years. The *Food Microbiology Research Conference*, organized in 1972 has had several sessions in rapid methods over the past 20 years of interactions of applied food microbiologists. Fung (1991a) compiled the "Food Microbiology Research Conference. 1972-1991. A Historical Review and Directory" to document all the topics of the 13 conferences, as well as listing names of participants of this unique conference series.

Many books and monographs on the general subject of Rapid Methods and Automation have appeared in recent years: Palmer and LeQuesne (1976), Mitruka (1976), Sharpe (1980), Eden and Eden (1984), Barry and Houghton (1986), Pierson and Stern (1986), Beerens (1987), Felix (1987), Hurst and Mortimer (1987), Erickson and Fung (1988), Houghton *et al.* (1988), Sharpe and Peterkin (1988), Adams and Hope (1989), Baltes (1989), Bills and Kung (1990), Fox *et al.* (1990), Fung and Matthews (1991), and Rossmoore (1991).

Another important development is the initiation of "hands-on" workshops concerning these rapid methods. Some of the workshops were one or two days with

lectures and limited demonstrations and "hands-on" experiences of various systems. The most comprehensive program was developed by the author in 1981. The program lasted 8 days. The first workshop had 16 participants from several countries. In 1990, the workshop celebrated its 10th anniversary with the addition of a mini-symposium. Through the years, the workshop has attracted more than 800 scientists from 30 countries and 40 states to participate in the activities. The 12th workshop will be held in July 1992 at Kansas State University, Manhattan, Kansas. A cumulative directory of all past participants of this workshop program and systems studied was issued by Fung (1991b). Similar symposia and workshops also were held on a smaller scale in Singapore; Taiwan; Malaysia; Australia; UK; Japan; River Falls, Wisconsin. Many more are planned in the coming year, such as those held in Taiwan in March and Guelph and Toronto, Canada, in May 1992.

General review articles on this subject were published by Goldschmidt and Fung (1979); Fung and Cox (1981); Cox *et al.* (1984, 1987); Fung (1980, 1984, 1989, 1991c); Fung *et al.* (1984, 1987, 1988, 1989); and Hartman *et al.* (1992). Some reviews were written in German (Fung 1975a,b), French (Fung and Lahellec 1987), Japanese (Fung 1990), modernized Chinese (Fung and Liang 1988b), and traditional Chinese (Fung and Chen 1992).

General themes of these publications on rapid methods can be grouped as follows:

Improvement in Sampling and Sample Preparations

Stomacher instrument. Emswiler *et al.* (1977); Sharpe and Jackson (1975)
 Gravimeter Diluter. Manninen and Fung (1992a)
 Surface sampling. Fung *et al.* (1980); Lee and Fung (1986)

Improvements of Conventional Procedures and Development of Diagnostic Kits

Miniaturized methods for:

Bacteria. Fung (1969); Fung and Hartman (1975); Fung and Miller (1970, 1972); Fung and Petrishko (1973); Lee *et al.* (1982, 1985).

Yeast and Mold. Lin and Fung (1985); Lin *et al.* (1984); Fung and Liang (1990).
 Motility enrichment oxyrase system. Yu and Fung (1991a,b, 1992).

Organic dye culture media. Fung and Miller (1973); Fung and Neimaic (1977);
 Chein and Fung (1990); Lin and Fung (1987); Fung and Liang (1988a);
 Hart *et al.* (1991); Goldschmidt *et al.* (1991).

Double tube anaerobic method. Fung and Lee (1981); Ali *et al.* (1991); Ali and Fung (1991, 1992)

Diagnostic kits: API, Enterotube, R/B, Minitex, Spectrum 10, MicroID, IDS.
 Cox *et al.* (1977, 1984, 1985, 1987).

Modifications of Viable Cell Count Procedure

Spiral system, Isogrid, Petrifilm, Redigel. Fung *et al.* (1987); Chain and Fung (1991); Manninen *et al.* (1991); Fung and Chain (1992).

DEFT test. Pettipher (1986).

Developments of New Methods for Estimation of Microbial Populations

ATP, Electrical Impedance, Conductance, Radiometry, Microcalorimetry,
 Limulus amoebocyte lysate, catalase, reflectance colorimetry. Fung (1991c);
 Manninen and Fung (1992b).

Identification of Microbes by Sophisticated Instruments and Procedures

Vitek, Sensititre, ELISA, DNA-RNA probes, 1-2 Test, protein profiling, cell wall components, magnetic beads, Polymerase Chain Reaction, bacterial ice nucleation test, etc. Fung (1989, 1991c).

KEY ISSUES AND CONCERNS

The key issues and concerns in the development of rapid methods are summarized in the following outline.

What is "Rapid"?

- A. Speed in Reaction
 - Instantaneous, seconds, minutes, hours
 - versus
 - Hours, days, weeks
- B. Numbers of Samples Per Operation
 - 10, 50, 96, or more
 - versus
 - one at a time

What is "Automation"?

- A. Manual: By humans, one or more sample at a time
- B. Semi-automation: By humans, with the aid of some form of instrumentation
- C. Complete Automation: By robots

Major Developments

- A. Diagnostic Tests for Liquid, Semi-Solid, and Solid Media
 1. Large tubes: One reaction per tube
 2. Large tubes: Multiple reactions per tube
 3. Small tubes: One reaction per tube
 4. Small tubes: Multiple reactions per tube
 5. Wells in a tray of different configurations:
 - a. One type of reaction for many organisms per tray
 - b. Many types of reactions for one type of organism per tray
 - c. Many types of reactions for a few organisms per tray
 6. Diagnostic kits
- B. Inoculations into Diagnostic Tests
 1. One inoculation per tube
 2. Multiple inoculations (manually or by instruments)
 - a. liquid in a tray
 - b. solid in agar
 - c. agar in multiple compartments
 - d. liquid dispensing to multiple wells
- C. Automated Instruments for Monitoring
 1. Cell mass
 2. Cell components
 3. Cell metabolites
 4. Cell activities
- D. Development of Serological and Immunological Techniques
 1. Immunoblotting
 2. Electrophoresis
 3. RIA
 4. ELISA
- E. Development of Genetic Techniques
 1. DNA probes, RNA probes
 2. Polymerase Chain Reaction

Concepts Involving the Living Cell

- A. Living cell versus dead cell
- B. Growing cell versus nongrowing cell
- C. Meaning of viable cell count
- D. Correlation between total count and other parameters
- E. Amplification of cells
- F. Concentration of cells
- G. Signal versus background
- H. Sensitivity versus detection time

SUMMARY

After being involved in the field of rapid methods and automation in microbiology for 25 years, this author summarizes his ideas concerning the attributes for an ideal automated microbiology assay system as follows:

Attributes for an Idea Automated Microbiology Assay System

1. Accuracy for the intended purposes
 - sensitivity; minimal detectable limits
 - specificity of test system
 - versatility; potential applications
 - comparison to referenced methods
2. Speed in productivity
 - in obtaining results
 - in number of samples processed per run; per day
3. Cost
 - initial, per test, reagents, others
4. Acceptability
 - by scientific community
 - by regulatory agencies
5. Simplicity of operation
 - sample preparation
 - operation of test equipment
 - computer versatility
6. Training
 - on site; how long
 - quality of training personnel
7. Reagents
 - reagent preparation-stability-availability-consistency
8. Company reputation
9. Technical service
 - speed and availability
 - cost and scope of technical background
10. Utility and space requirement

Although the author marvels at the great changes in rapid methods and automation in microbiology in the past 25 years, he is still of the opinion that all students of microbiology should thoroughly master the basic conventional methods first (i.e., Gram stain, aseptic techniques, pure culture isolation, test tube methods,

standard plate count method, skill in microscope utilization, etc.) before tackling all these newer concepts, instruments, and sophisticated techniques. He likes to end a discussion on rapid methods and automation in microbiology by saying "The best rapid method is still your own brain."

In conclusion, rapid methods and automation in microbiology as a field of study is relatively new. The potential is great and many exciting developments will certainly unfold in the coming years.

REFERENCES

- ADAMS, M.R. and HOPE, C.F.A. 1989. *Rapid Methods in Food Microbiology*, Vol. 26, Elsevier Science Publishing Co., New York.
- ALI, M.S. and FUNG, D.Y.C. 1991. Occurrence of *Clostridium perfringens* in ground beef and ground turkey evaluated by three methods. *J. Food Safety* 11, 197-203.
- ALI, M.S. and FUNG, D.Y.C. 1992. Detection of enterotoxigenic *Clostridium perfringens* in ground beef. *J. Rapid Methods and Automation in Microbiol.* 1(2) (In Press.), 165-170.
- ALI, M.S. FUNG, D.Y.C., and KASTNER, C.L. 1991. Comparison of rapid methods for isolation and enumeration of *Clostridium perfringens* in meat. *J. Food Sci.* 56, 367-370.
- BALOWS, A., TILTON, R.C. and TURANO, A. 1989. 5th International Symposium: *Rapid Methods and Automation in Microbiology and Immunology*, Tipografia Queriniana, Brescia, Italy.
- BALTES, W. 1989. *Rapid Methods for Analysis of Foods and Food Raw Material*, Technomic Publishing Co., Lancaster, PA.
- BARRY, S.B. and HOUGHTON, D.R. 1986. *Biodeterioration* 6, C.A.B. International, Slough, U.K.
- BEERENS, H. 1987. *Sciences Des Aliments*, Lavoisier Abonnements, Paris.
- BILLS, D. and KUNG, S. 1990. *Biotechnology and Food Safety*, Butterworth-Heinemann, Stoneham, MA.
- CHAIN, V.S. and FUNG, D.Y.C. 1991. Comparative analysis of Redigel, Petrifilm, Isogrid, Spiral plating systems and the standard plate count method for the evaluation of mesophiles from selected foods. *J. Food Protect.* 54, 208-211.
- CHEIN, S.P. and FUNG, D.Y.C. 1990. Acriflavine violet red bile agar for the isolation and enumeration of *Klebsiella pneumoniae*. *Food Microbiol.* 7, 73-79.
- COX, N.A., BAILEY, J.S., FUNG, D.Y.C. and HARTMAN, P.A. 1987. Rapid methods for the detection and identification of microorganisms in foods. In *Food Protection Technology*, (C.W. Felix, ed.) pp. 125-131, Lewis Publishers, Chelsea, MI.
- COX, N.A., FUNG, D.Y.C., GOLDSCHMIDT, M.C., BAILEY, J.S. and THOMSON, J.E. 1984. Selecting a miniaturized system for identification of Enterobacteriaceae. *J. Food Protect.* 47, 74-77.
- COX, N.A., MCHAN, F. and FUNG, D.Y.C. 1977. Commercially available minikits for the identification for Enterobacteriaceae: A review. *J. Food Protect.* 40, 866-872.
- COX, N.A., VAN WART, M., BAILEY, J.S. and THOMSON, J.E. 1985. Identification of Enterobacteriaceae from foods with Spectrum 10. *J. Food Protect.* 48, 76-79.
- EDEN, R.F. and EDEN, G. 1984. *Impedance Microbiology*, John Wiley & Sons, New York.
- EMSWILER, B.S., PIERSON, C.J. and KOTULA, A.W. 1977. Stomaching vs blending. *Food Technol.* 31(10), 40-42.
- ERICKSON, L.E. and FUNG, D.Y.C. 1988. *Handbook on Anaerobic Fermentations*, Marcel Dekker, New York.
- FELIX, C.W. 1987. *Food Protection Technology*, Lewis Publishers, New York.
- FOX, A., MORGAN, S.L., LARSON, L. and ODHAM, G. 1990. *Analytical Microbiology Methods: Chromatograph and Mass Spectrometry*, Plenum Press, New York.
- FUNG, D.Y.C. 1969. Rapid Methods for Determining Staphylococcal Toxins and *Salmonellae* Associated with Poultry Products. Ph. D. dissertation. Iowa State University Library, Ames, IA.
- FUNG, D.Y.C. 1971. Miniaturization of microbiological methods for rapid characterization of microorganisms. Symposium on rapid methods for detection and characterization of microorganisms. Annual Meeting American Society for Microbiology, May, Minneapolis, MN.
- FUNG, D.Y.C. 1975a. Miniaturisierte Mikrobiologische Methoden zur Charakterisierung von Bakterien. *Arch. Lebensmittel-hygiene* 26, 18-20. (In German)
- FUNG, D.Y.C. 1975b. Halbautomatisierte Techniken zur Zählung lebender Bakterien und ein automatisiertes Verfahren. *Arch. Lebensmittel-hygiene.* 26, 27-29. (In German)
- FUNG, D.Y.C. 1980. Teaching automation and rapid methods in food microbiology. *J. Food Protect.* 43, 733-735.
- FUNG, D.Y.C. 1984. Rapid methods for determining the bacteriology quality of red meats. *J. Environ. Health* 46, 226-228.
- FUNG, D.Y.C. 1989. Rapid methods and automation in meat microbiology: A review. *Food Australia* 41(12), 1094-1097.
- FUNG, D.Y.C. 1990. Rapid methods in food microbiology. Proceedings of Tokyo and Kobe Rapid Methods Symposium, Mar. 7-9. (In Japanese)

- FUNG, D.Y.C. 1991a. *Food Microbiology Research Conference. 1972-1992. A Historical Review and Directory*, Kansas State University, Manhattan, KS.
- FUNG, D.Y.C. 1991b. *Participant Directory for 1981#1991. Rapid Methods and Automation in Microbiology*, Kansas State University, Manhattan, KS.
- FUNG, D.Y.C. 1991c. Rapid methods and automation for food microbiology. In *Instrumental Methods for Quality Assurance in Foods*, (D.Y.C. Fung and R.F. Matthews, eds.), Marcel Dekker, New York.
- FUNG, D.Y.C. and CHAIN, V.S. 1992. Comparative analysis of Redigel and aerobic plate counts methods for viable cell count of selected foods. *Food Microbiol.* 8, 299-301.
- FUNG, D.Y.C. and CHEN, H.L.W. 1992. Rapid methods in food microbiology. *Food Ind. Res. Dev. Inst. J.* 24(1), 8-14. (In Chinese)
- FUNG, D.Y.C. and COX, N.A. 1981. Rapid identification systems in the food industry: Present and future. *J. Food Protect.* 44, 877-880.
- FUNG, D.Y.C., COX, N.A. and BAILEY, J.S. 1988. Rapid methods and automation in microbiology. *Dairy Food Sanit.* 8(6), 292-296.
- FUNG, D.Y.C., COX, N.A., GOLDSCHMIDT, M.C. and BAILEY, J.S. 1989. Rapid methods and automation: A survey of professional microbiologists. *J. Food Protect.* 52, 65-68.
- FUNG, D.Y.C., GOLDSCHMIDT, M.C. and COX, N.A. 1984. Evaluation of bacterial diagnostic kits and systems at an instructional workshop. *J. Food Protect.* 47, 68-73.
- FUNG, D.Y.C., HART, R.A. and CHAIN, V. 1987. Rapid methods and automated procedures for microbiological evaluation of seafood. In *Seafood Quality Determination*, (D.E. Kramer and J. Liston, ed.) pp. 247-253, Elsevier Science Publisher, Amsterdam.
- FUNG, D.Y.C. and HARTMAN, P.A. 1975. Miniaturized microbiological techniques for rapid characterization of bacteria. In *New Approaches to the Identification of Microorganisms*, (G.C. Heden and T. Illeni, ed.) Chap. 21, John Wiley & Sons, New York.
- FUNG, D.Y.C. and LAHELLEC, C. 1987. Techniques rapides et automationen microbiologie alimentaire. *Sci. Aliments* 7(1987) n° hors-serie VII, 5-11. (In French)
- FUNG, D.Y.C. and LEE, C.M. 1981. Double tube anaerobic bacterial cultivation system. *Food Sci.* 7, 209-213. (In Chinese)
- FUNG, D.Y.C., LEE, C.Y. and KASTNER, C.L. 1980. Adhesive tape method for estimation of microbial load on meat surface. *J. Food Protect.* 43, 295-297.
- FUNG, D.Y.C. and LIANG, C. 1988a. Un nouveau milieu gelose fluorescent pour l'isolement de *Candida albicans*. *Technique de laboratoire Bull. Lab. Vet. No. 29/30. Murs/Jain.* (In French)
- FUNG, D.Y.C. and LIANG, C. 1988b. Rapid methods and automation in

- microbiology for biomass estimation. *Microbiology* 15(9), 173-177. (In Chinese)
- FUNG, D.Y.C. and LIANG, C. 1990. Critical review of isolation, detection and identification of yeasts from meat products. *CRC Crit. Rev. Foods Nutr.* 29(5), 341-379.
- FUNG, D.Y.C. and MATTHEW, R.F. 1991. *Instrumental Method for Quality Assurance and Research in Foods*, Marcel Dekker, New York.
- FUNG, D.Y.C. and MILLER, R.D., 1970. Rapid procedure for the detection of acid and gas production by bacterial cultures. *Appl. Microbiol.* 20, 527-528.
- FUNG, D.Y.C. and MILLER, R.D., 1972. Miniaturized techniques for IMVIC tests. *J. Milk Food Technol.* 35, 328-329.
- FUNG, D.Y.C. and MILLER, R.D., 1973. Rapid screening procedure for testing the effects of dyes on bacterial growth. *Appl. Microbiol.* 25, 793-799.
- FUNG, D.Y.C. and NEIMEIC, M. 1977. Acriflavine violet red bile medium for the isolation of *Klebsiella*. *Lab Health Sci.* 14, 273-278.
- FUNG, D.Y.C. and PETRISHKO, D.T. 1973. Capillary tube catalase test. *Appl. Microbiol.* 26, 631-632.
- GALTEAU, M.M., HENNY, J. and SIEST, G. 1989. *Biologie Perspective*, John Libbey & Co., London.
- GOLDSCHMIDT, M.C. 1970. Instrumentation for microbiology: Horizons unlimited. In *Rapid Diagnosis Methods in Medical Microbiology*, (C.D. Graber, ed.) Chap. 24, William & Wilkins Co., Baltimore.
- GOLDSCHMIDT, M.C. 1971. Biochemical and biophysical parameters used in rapid detection and characterization of microorganisms. Symposium in Rapid Methods for Detection and Characterization of Microorganisms. Annual Meeting American Society for Microbiology, May, Minneapolis, MN.
- GOLDSCHMIDT, M.C. and FUNG, D.Y.C. 1979. Automated and new instrumentation for microbiological analysis. *Food Technol.* 32, 63-70.
- GOLDSCHMIDT, M.C., FUNG, D.Y.C., GRANT, R., WHITE, J. and BROWN, T. 1991. New aniline blue dye medium for rapid identification and isolation of *Candida albicans*. *J. Clin. Microbiol.* 28(6), 1095-1099.
- GRABER, C.D. 1970. *Rapid Diagnosis Methods in Microbiology*, Williams & Wilkins, Co., Baltimore.
- HABERMEHL, K.O. 1984. 4th International Symposium: *Rapid Methods and Automation in Microbiology and Immunology*, Springer-Verlag, Berlin.
- HART, R.A., MO, O., BORJUS, F. and FUNG, D.Y.C. 1991. Comparative analysis of Trypan blue agar and congo red agar for the enumeration of yeast and mold using HGMF System. *J. Food Safety* 11, 227-230.
- HARTMAN, P.A. 1968. *Miniaturized Microbiological Methods*, Academic Press, New York.
- HARTMAN, P.A. et al. 1992. Rapid methods and automation. *Compendium of*

- Methods for the Microbiological Examination of Foods*, 3rd Ed., (C. Vanderzant and D.F. Splittstoesser, eds.) Chap. X, Am. Publ. Health Association, Arlington, VA.
- HEDEN, C.G. and ILLENI, T. 1975a. 1st International Symposium: *Automation in Microbiology and Immunology*, John Wiley & Sons, New York.
- HEDEN, C.G. and ILLENI, T. 1975b. 1st International Symposium: *New Approaches to the Identification of Microorganisms*, John Wiley & Sons, New York.
- HOUGHTON, D.R., SMITH, R.N. and EGGINS, H.O.W. 1988. *Biodeterioration 7*, Elsevier Science Publisher, New York.
- HUNG, C.B. 1987. *Proceedings of Workshop on the Science and Technology of Food and Drug Safety*, Drug and Food Bureau, Taipei, Taiwan.
- HURST, W.J. and MORTIMER, J.W. 1987. *Laboratory Robotics*, V.C.H. Publishers, New York.
- KRAMER, D. and LISTON, J. 1987. *Seafood Quality Determination*, Elsevier Science Publishers, New York.
- LACLERC, H. 1983. *Gram Negative Bacteria of Medical and Public Health Importance*, Editions INSERM, Paris.
- LAHELLEC, C. 1983. *Quality of Poultry Meat*, Ministere de l'Agriculture, Ploufragan, France.
- LEE, C.Y., FUNG, D.Y.C. and KASTNER, C.L. 1982. Computer-assisted identification of bacteria on hot-boned and conventionally processed beef. *J. Food Sci.* 47, 363-367.
- LEE, C.Y., FUNG, D.Y.C. and KASTNER, C.L. 1985. Computer-assisted identification of microflora on hot-boned and conventionally processed beef: Effect of moderate and slow initial chilling rate. *J. Food Sci.* 50, 553-567.
- LEE, J.Y. and FUNG, D.Y.C. 1986. Surface sampling techniques for bacteriology. *J. Environ. Health* 48, 200-205.
- LIN, C.C.S. and FUNG, D.Y.C. 1985. Effect of dyes on the growth of food yeast. *J. Food Sci.* 50, 241-244.
- LIN, C.C.S. and FUNG, D.Y.C. 1987. Critical review of conventional and rapid methods for yeast identification. *CRC Crit. Rev. Microbiol.* 14(4), 273-289.
- LIN, C.C.S., FUNG, D.Y.C. and ENTIS, P. 1984. Growth of yeast and mold on Trypan Blue agar in conjunction with the ISOGRID system. *Can. J. Microbiol.* 30, 1405-1407.
- MANNINEN, M.T. and FUNG, D.Y.C. 1992a. Use of Gravimetric diluter in microbiological work. *J. Food Protect.* 55, 59-61.
- MANNINEN, M.T. and FUNG, D.Y.C. 1992b. Estimation of microbial members from pure bacterial cultures and from minced beef samples by reflectance colorimetry with Omnispec 4000. *J. Rapid Methods and Automation in Microbiol.* 1(1), pp. 41-58.

- MANNINEN, M.T., FUNG, D.Y.C. and HART, R.A. 1991. Spiral System and laser counter for enumerations of microorganism. *J. Food Safety* 11, 177-187.
- MITRUKA, B.M. 1976. *Methods of Detection and Identification*, CRC Press, Boca Raton, FL.
- NEWSOM, S.W.B. 1976. 2nd International Symposium: *Rapid Methods and Automation in Microbiology*, Learned Information (Europe) Ltd., Oxford.
- PALMER, W. and LEQUESNE, S. 1976. *Rapid Methods and Automations in Microbiology and Immunology: A Bibliography*, Information Retrieval Ltd, Washington, DC.
- PETTIPHER, G.L. 1986. Review: The direct epifluorescent filter technique. *J. Food Technol.* 21, 535-546.
- PIERSON, M.D. and STERN, N.J. 1986. *Foodborne Microorganisms and Their Toxins: Developing Methodology*, Marcel Dekker, New York.
- RICHARDSON, G.H. 1972. Automation in the dairy laboratory. *J. Milk Food Technol.* 35(5), 279-287.
- ROSSMOORE, H.W. 1991. *Biodeterioration and Biodegradation 8*. Elsevier Applied Science, New York.
- SCHINDLER, J. 1980. *Conference on Taxonomy, Computer Identification of Bacteria and Diagnostic Methods*, May 2-9, Liblice Castle, Czechoslovakia.
- SHARPE, A.N. 1980. *Food Microbiology*, Charles C. Thomas Publishing Co., Springfield, IL.
- SHARPE, A.N. and CLARK, D.S. 1978. *Mechanizing Microbiology*, Charles C. Thomas Publishing Co., Springfield, IL.
- SHARPE, A.N. and JACKSON, A.K. 1975. Automation requirements in microbiological quality control of foods. (C.G. Heden and T. Illen, eds.) *Automation in Microbiology and Immunology*, John Wiley & Sons, New York.
- SHARPE, A.N. and PETERKIN, P.I. 1988. *Membrane Filter Food Microbiology*, John Wiley & Sons, New York.
- SKERMAN, V.B.D. 1969. *Abstracts of Microbiological Methods*, John Wiley & Sons, New York.
- TILTON, R.C. 1982. 3rd International Symposium: *Rapid Methods and Automation in Microbiology*, American Society for Microbiology, Washington, DC.
- TOLLE, A. and HEESCHEN, W. 1974. 1st International Symposium: *Rapid Methods and Automation in Food Microbiology*, Sept. 16-20, Milk Research Institute, Kiel, Germany.
- TROTMAN, R.E. 1973. The philosophy of the application of automatic methods to hospital diagnostic bacteriology. *Biomed. Eng.* 8(12), 519-521.
- VAHERI, A., TILTON, R.C. and BALOWS, A. 1991. 6th International Symposium: *Rapid Methods and Automation in Microbiology and Immunology*, Springer-Verlag, New York.
- YU, L.S.L. and FUNG, D.Y.C. 1991a. Oxyrase enzyme and motility enrich-

ment. Fung-Yu tube for rapid detection of *Listeria monocytogenes* and *Listeria* spp. J. Food Safety 11, 149-162.

YU, L.S.L. and FUNG, D.Y.C. 1991b. Effect of oxyrase enzyme in *Listeria monocytogenes* and other anaerobes. J. Food Safety 11, 163-176.

YU, L.S.L. and FUNG, D.Y.C. 1992. Comparison of selected methods with Fung-Yu procedures for detecting *Listeria monocytogenes* and other *Listeria* spp. in meats. J. Food Protect. (In Press.) 55, 349-355.